

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 002

Vol. flask, 20 ml, NS 10, amber glass, A, conf. cert.

Batch: 36-23


Serial:

Measuring Nr. 64421

Batch size:

1198

Volume:	20,0000 ml
Tolerance:	± 0,0400 ml
Average value:	19,9935 ml
Deviation:	0,0033 ml
Precision (CV):	0,0163 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

18.09.2023

Balance:

Sartorius -
LE244S-OCE

DE-M