

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 003

Vol. flasks, 25 ml, NS 10, amber glass, A, conf. cert.


Batch: 50-22

Serial:

Measuring Nr. 60837

Batch size: 1199

Volume:	25,0000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	24,9962 ml
Deviation:	0,0077 ml
Precision (CV):	0,0306 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	77
Date:	05.01.2023
Balance:	Sartorius - LE244S-OCE

DE-M