

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 005 USP

Vol. flasks, 50 ml, NS 12, amber glass, A, conf. cert.

Batch: 14-23


Serial:

Measuring Nr. 62508

Batch size:

431

Volume:	50,0000 ml
Tolerance:	± 0,0500 ml
Average value:	50,0224 ml
Deviation:	0,0058 ml
Precision (CV):	0,0115 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

11.04.2023

Balance:

Sartorius -
LE244S-OCE

DE-M