

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 670 012

**Vol. flasks, 10 ml, NS 10, amber glass, A, conf. cert.**


**Batch:** 24-24

**Serial:**

**Measuring Nr.** 67694

**Batch size:** 497

<b>Volume:</b>	10,0000 ml
<b>Tolerance:</b>	± 0,0400 ml
<b>Average value:</b>	9,9961 ml
<b>Deviation:</b>	0,0029 ml
<b>Precision (CV):</b>	0,0288 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	17.06.2024
<b>Balance:</b>	Sartorius - LE244S-OCE

DE-M