

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 013 BL PP

Vol.flask, 20 ml, NS 12, amber glass, blue grad.,A, conf.


Batch: 06-25

Serial:

Measuring Nr. 72026

Batch size: 194

Volume:	20,000 ml
Tolerance:	$\pm 0,0600$ ml
Average value:	19,9987 ml
Deviation:	0,0028 ml
Precision (CV):	0,0141 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	20.02.2025
Balance:	Sartorius - LE244S-OCE

DE-M