

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 016

Vol. flask, 15 ml, NS 10, amber glass, A, conf. cert.

Batch: 49-19


Serial:

Measuring Nr. 35094

Batch size:

280

Volume:	15,0000 ml
Tolerance:	$\pm 0,0400$ ml
Average value:	14,9977 ml
Deviation:	0,0078 ml
Precision (CV):	0,0521 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

9

Date:

09.12.2019

Balance:

Sartorius -
LE244S-OCE

DE-M