

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 018

Vol. flasks, 25 ml, NS 14, amber glass, A, conf. cert.


Batch: 44-18

Serial:

Measuring Nr. 25712

Batch size: 148

Volume:	25,0000 ml
Tolerance:	$\pm 0,0600$ ml
Average value:	24,9886 ml
Deviation:	0,0118 ml
Precision (CV):	0,0471 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	20.11.2018
Balance:	Sartorius - LE244S-OCE

