

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 020

Vol. flask, 200 ml, NS 14, amber glass, A, conf. cert.

Batch: 31-20


Serial:

Measuring Nr. 39441

Batch size:

1199

Volume:	200,000 ml
Tolerance:	$\pm 0,1500$ ml
Average value:	200,0089 ml
Deviation:	0,0109 ml
Precision (CV):	0,0054 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

06.08.2020

Balance:

Sartorius -
MSE224S-0CE
-DU

DE-M