

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 050

Vol. flasks, 500 ml, NS 19, amber glass, A, conf. cert.


Batch: 35-18

Serial:

Measuring Nr. 23802

Batch size: 500

Volume:	500,000 ml
Tolerance:	$\pm 0,2500$ ml
Average value:	499,9582 ml
Deviation:	0,0627 ml
Precision (CV):	0,0125 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	05.09.2018
Balance:	Sartorius - LA 1200 S

DE-M