

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 670 050

**Vol. flasks, 500 ml, NS 19, amber glass, A, conf. cert.**

**Batch:** 19-19


**Serial:**

**Measuring Nr.** 29893

**Batch size:**

323

<b>Volume:</b>	500,000 ml
<b>Tolerance:</b>	$\pm 0,2500$ ml
<b>Average value:</b>	499,9569 ml
<b>Deviation:</b>	0,0491 ml
<b>Precision (CV):</b>	0,0098 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

15.05.2019

**Balance:**

Sartorius - LC  
6200 S

DE-M