

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 150

Vol. flasks, 150 ml, NS 14, amber glass, A, conf. cert.


Batch: 25-18

Serial:

Measuring Nr. 22404

Batch size: 280

Volume:	150,000 ml
Tolerance:	$\pm 0,1500$ ml
Average value:	149,9780 ml
Deviation:	0,0244 ml
Precision (CV):	0,0163 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	44
Date:	20.07.2018
Balance:	Sartorius - LE244S-OCE

DE-M