

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 670 500

Vol. flask, 5000 ml, NS 34, amber glass, A, conf. cert.


Batch: 17-19

Serial:

Measuring Nr. 29709

Batch size: 15

Volume:	5000,0000 ml
Tolerance:	$\pm 1,2000$ ml
Average value:	4999,6784 ml
Deviation:	0,4224 ml
Precision (CV):	0,0084 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	08.05.2019
Balance:	Sartorius - MSE14202S- OCE-D0

DE-M