

Certificate of performance



Article: 3 674 0382

Control volumetric flask, 368-382 ml, A, conf. cert.




Batch: 18-21

Serial:

Measuring Nr.: 46552

Batch size: 4

Volume:	370,000 ml	375,000 ml	380,000 ml
Tolerance:	± 0,2500 ml	± 0,2500 ml	± 0,2500 ml
Average value:	369,9128 ml	374,9984 ml	380,1008 ml
Deviation:	0,0586 ml	0,0673 ml	0,0652 ml
Precision (CV):	0,0158 %	0,0179 %	0,0172 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 1042

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	07.05.2021
Balance:	Sartorius - LA 1200 S

