

Certificate of performance



Article: 3 510 101

Grad. pipettes, 1:0,01 ml, AS, conf. cert.




Batch: 39-17

Serial:

Measuring Nr. 14908

Batch size: 3200

Volume:	1,0000 ml	0,6000 ml	0,2000 ml
Tolerance:	± 0,0070 ml	± 0,0070 ml	± 0,0070 ml
Average value:	1,0011 ml	0,6005 ml	0,2005 ml
Deviation:	0,0023 ml	0,0022 ml	0,0012 ml
Precision (CV):	0,2290 %	0,3662 %	0,5997 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	16.10.2017
Balance:	Sartorius - LE244S-OCE

