

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 510 202 USP

Grad. pipettes, 2:0, 0,2 ml, USP, AS, conf. cert.




Batch: 13-22

Serial:

Measuring Nr. 54281

Batch size: 150

Volume:	2,0000 ml	1,2000 ml	0,4000 ml
Tolerance:	± 0,0100 ml	± 0,0100 ml	± 0,0100 ml
Average value:	2,0021 ml	1,2026 ml	0,4020 ml
Deviation:	0,0024 ml	0,0027 ml	0,0015 ml
Precision (CV):	0,1180 %	0,2262 %	0,3833 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	07.04.2022
Balance:	Sartorius - LE244S-OCE

DE-M