

Certificate of performance



Article: 3 510 505 BL

Grad. pipettes, 5:0,05 ml, blue grad., AS, conf. cert.




Batch: 09-22

Serial:

Measuring Nr. 52489

Batch size: 500

Volume:	5,0000 ml	3,0000 ml	1,0000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	4,9912 ml	3,0045 ml	1,0030 ml
Deviation:	0,0054 ml	0,0058 ml	0,0036 ml
Precision (CV):	0,1070 %	0,1943 %	0,3637 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	28.02.2022
Balance:	Sartorius - LE244S-OCE

