

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 510 710

Graduated pipettes, 10:0,1 ml, AS, conf. cert.




Batch: 41-20

Serial:

Measuring Nr. 41504

Batch size: 4318

Volume:	10,0000 ml	6,0000 ml	2,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	9,9948 ml	5,9960 ml	2,0026 ml
Deviation:	0,0077 ml	0,0085 ml	0,0058 ml
Precision (CV):	0,0768 %	0,1409 %	0,2919 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	20.10.2020
Balance:	Sartorius - LE244S-OCE

