

Certificate of performance



Article: 3 510 810

Graduated pipettes, 20:0,1 ml, AS, conf. cert.




Batch: 24-24

Serial:

Measuring Nr. 67713

Batch size: 1200

Volume:	20,000 ml	12,000 ml	4,000 ml
Tolerance:	± 0,1000 ml	± 0,1000 ml	± 0,1000 ml
Average value:	19,9948 ml	12,0377 ml	4,0250 ml
Deviation:	0,0214 ml	0,0181 ml	0,0132 ml
Precision (CV):	0,1068 %	0,1509 %	0,3294 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	19.06.2024
Balance:	Sartorius - LE244S-OCE

