

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 515 710

Grad. pipettes, 10:0,1 ml, AS, conf. cert.




Batch: 33-18

Serial:

Measuring Nr.: 23852

Batch size: 1200

Volume:	10,0000 ml	6,0000 ml	2,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	9,9888 ml	5,9893 ml	1,9969 ml
Deviation:	0,0130 ml	0,0086 ml	0,0035 ml
Precision (CV):	0,1295 %	0,1438 %	0,1725 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	44
Date:	07.09.2018
Balance:	Sartorius - LE244S-OCE

DE-M