

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 515 710

Grad. pipettes, 10:0,1 ml, AS, conf. cert.




Batch: 40-21

Serial:

Measuring Nr. 49645

Batch size: 1500

Volume:	10,000 ml	6,0000 ml	2,0000 ml
Tolerance:	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
Average value:	9,9790 ml	5,9896 ml	1,9989 ml
Deviation:	0,0121 ml	0,0091 ml	0,0060 ml
Precision (CV):	0,1211 %	0,1517 %	0,3005 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	11.10.2021
Balance:	Sartorius - LE244S-OCE

