

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 005

**Volumetric pipettes, 0,5 ml, AS, conf. cert.**

**Batch:** 37-17


**Serial:**

**Measuring Nr.** 14585

**Batch size:**

500

|                        |                  |
|------------------------|------------------|
| <b>Volume:</b>         | <b>0,5000 ml</b> |
| <b>Tolerance:</b>      | $\pm 0,0050$ ml  |
| <b>Average value:</b>  | 0,4997 ml        |
| <b>Deviation:</b>      | 0,0008 ml        |
| <b>Precision (CV):</b> | 0,1649 %         |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

4

**Date:**

05.10.2017

**Balance:**

Sartorius -  
LE244S-OCE

DE-M