

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 010

**Volumetric pipettes, 1 ml, AS, conf. cert.**

**Batch:** 46-21


**Serial:**

**Measuring Nr.** 50884

**Batch size:**

3191

<b>Volume:</b>	1,0000 ml
<b>Tolerance:</b>	$\pm 0,0080$ ml
<b>Average value:</b>	0,9990 ml
<b>Deviation:</b>	0,0014 ml
<b>Precision (CV):</b>	0,1435 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

29.11.2021

**Balance:**

Sartorius -  
LE244S-OCE

DE-M