

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 020 USP

**Volumetric pipettes, 2 ml, USP, AS, conf. cert.**

**Batch:** 48-23


**Serial:**

**Measuring Nr.** 65419

**Batch size:**

280

<b>Volume:</b>	2,0000 ml
<b>Tolerance:</b>	± 0,0060 ml
<b>Average value:</b>	1,9995 ml
<b>Deviation:</b>	0,0013 ml
<b>Precision (CV):</b>	0,0673 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

04.12.2023

**Balance:**

Sartorius -  
LE244S-OCE

DE-M