

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 020

Volumetric pipettes, 2 ml, AS, conf. cert.

Batch: 11-21


Serial:

Measuring Nr. 47105

Batch size:

2398

Volume:	2,0000 ml
Tolerance:	$\pm 0,0100$ ml
Average value:	2,0001 ml
Deviation:	0,0016 ml
Precision (CV):	0,0778 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	08.06.2021
Balance:	Sartorius - LE244S-OCE

DE-M