

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 040

Volumetric pipettes, 4 ml, AS, conf. cert.


Batch: 34-18

Serial:

Measuring Nr. 23803

Batch size: 2395

Volume:	4,0000 ml
Tolerance:	$\pm 0,0150$ ml
Average value:	4,0011 ml
Deviation:	0,0036 ml
Precision (CV):	0,0892 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 44

Date: 06.09.2018

Balance: Sartorius -
LE244S-OCE

DE-M