

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 040

Volumetric pipettes, 4 ml, AS, conf. cert.

Batch: 04-23


Serial:

Measuring Nr. 61644

Batch size:

500

| | |
|------------------------|------------------|
| Volume: | 4,0000 ml |
| Tolerance: | $\pm 0,0150$ ml |
| Average value: | 4,0024 ml |
| Deviation: | 0,0030 ml |
| Precision (CV): | 0,0745 % |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

09.02.2023

Balance:

Sartorius -
LE244S-OCE

DE-M