

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 050 USP

Volumetric pipette, 5 ml, USP, AS, conf. cert.


Batch: 44-20

Serial:

Measuring Nr. 42584

Batch size: 280

Volume:	5,0000 ml
Tolerance:	$\pm 0,0100$ ml
Average value:	5,0001 ml
Deviation:	0,0016 ml
Precision (CV):	0,0317 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 16.11.2020

Balance: Sartorius -
LE244S-OCE

DE-M