

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 060

Volumetric pipettes, 6 ml, AS, conf. cert.


Batch: 19-20

Serial:

Measuring Nr. 38232

Batch size: 280

| | |
|------------------------|------------------|
| Volume: | 6,0000 ml |
| Tolerance: | $\pm 0,0150$ ml |
| Average value: | 5,9985 ml |
| Deviation: | 0,0024 ml |
| Precision (CV): | 0,0397 % |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

| | |
|------------------|-----------------------------------|
| Operator: | 7 |
| Date: | 27.05.2020 |
| Balance: | Sartorius - MSE224S-0CE -DU |

DE-M