

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 060

**Volumetric pipettes, 6 ml, AS, conf. cert.**

**Batch:** 19-22


**Serial:**

**Measuring Nr.** 56541

**Batch size:**

500

<b>Volume:</b>	6,0000 ml
<b>Tolerance:</b>	$\pm 0,0150$ ml
<b>Average value:</b>	6,0000 ml
<b>Deviation:</b>	0,0015 ml
<b>Precision (CV):</b>	0,0244 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	4
<b>Date:</b>	13.06.2022
<b>Balance:</b>	Sartorius - MSE224S-0CE -DU

DE-M