

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 070

Volumetric pipettes, 7 ml, AS, conf. cert.


Batch: 06-23

Serial:

Measuring Nr. 61875

Batch size: 496

Volume:	7,0000 ml
Tolerance:	$\pm 0,0150$ ml
Average value:	7,0006 ml
Deviation:	0,0029 ml
Precision (CV):	0,0409 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	06.03.2023
Balance:	Sartorius - LE244S-OCE

DE-M