

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 100 USP

Volumetric pipettes, 10 ml, AS, USP, conf. cert.


Batch: 09-18

Serial:

Measuring Nr. 19733

Batch size: 280

Volume:	10,0000 ml
Tolerance:	$\pm 0,0200$ ml
Average value:	10,0076 ml
Deviation:	0,0023 ml
Precision (CV):	0,0232 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	44
Date:	10.04.2018
Balance:	Sartorius - LE244S-OCE

DE-M