

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 100

**Volumetric pipettes, 10 ml, AS, conf. cert.**

**Batch:** 27-17


**Serial:**

**Measuring Nr.** 12055

**Batch size:**

3194

|                        |                   |
|------------------------|-------------------|
| <b>Volume:</b>         | <b>10,0000 ml</b> |
| <b>Tolerance:</b>      | $\pm 0,0200$ ml   |
| <b>Average value:</b>  | 10,0005 ml        |
| <b>Deviation:</b>      | 0,0045 ml         |
| <b>Precision (CV):</b> | 0,0452 %          |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

44

**Date:**

13.07.2017

**Balance:**

Sartorius -  
LE244S-OCE

DE-M