

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 150 BL

Volumetric pipettes, 15 ml, blue grad., AS, conf. cert.

Batch: 07-24


Serial:

Measuring Nr. 66059

Batch size:

150

Volume:	15,0000 ml
Tolerance:	$\pm 0,0200$ ml
Average value:	15,0076 ml
Deviation:	0,0032 ml
Precision (CV):	0,0212 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

13.02.2024

Balance:

Sartorius -
LE244S-OCE

DE-M