

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 150 USP

Volumetric pipette, 15 ml, USP, AS, conf. cert


Batch: 22-22

Serial:

Measuring Nr. 56849

Batch size: 90

Volume:	15,0000 ml
Tolerance:	$\pm 0,0300$ ml
Average value:	15,0011 ml
Deviation:	0,0032 ml
Precision (CV):	0,0213 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	29.06.2022
Balance:	Sartorius - LE244S-OCE

DE-M