

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 150

Volumetric pipettes, 15 ml, AS, conf. cert.

Batch: 11-23


Serial:

Measuring Nr. 62376

Batch size:

797

Volume:	15,0000 ml
Tolerance:	$\pm 0,0200$ ml
Average value:	15,0027 ml
Deviation:	0,0057 ml
Precision (CV):	0,0382 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

30.03.2023

Balance:

Sartorius -
LE244S-OCE

DE-M