

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 170

**Volumetric pipettes, 17 ml, AS, conf. cert**

**Batch:** 14-18


**Serial:**

**Measuring Nr.** 19651

**Batch size:**

60

<b>Volume:</b>	<b>17,0000 ml</b>
<b>Tolerance:</b>	$\pm 0,0300$ ml
<b>Average value:</b>	17,0072 ml
<b>Deviation:</b>	0,0071 ml
<b>Precision (CV):</b>	0,0418 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	44
<b>Date:</b>	09.04.2018
<b>Balance:</b>	Sartorius - LE244S-OCE

DE-M