

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 170

Volumetric pipettes, 17 ml, AS, conf. cert


Batch: 07-21

Serial:

Measuring Nr. 44667

Batch size: 49

Volume:	17,0000 ml
Tolerance:	$\pm 0,0300$ ml
Average value:	17,0032 ml
Deviation:	0,0058 ml
Precision (CV):	0,0343 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	4
Date:	26.02.2021
Balance:	Sartorius - MSE224S-0CE -DU

DE-M