

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 190 BL

Volumetric pipettes, 19 ml, blue grad., AS, KB


Batch: 44-19

Serial:

Measuring Nr. 34257

Batch size: 150

Volume:	19,0000 ml
Tolerance:	$\pm 0,0300$ ml
Average value:	19,0090 ml
Deviation:	0,0043 ml
Precision (CV):	0,0226 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	07.11.2019
Balance:	Sartorius - MSE224S-0CE -DU

DE-M