

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 200 BL

Volumetric pipettes, 20 ml, blue grad., AS, conf. cert.


Batch: 26-24

Serial:

Measuring Nr. 67934

Batch size: 493

Volume:	20,0000 ml
Tolerance:	$\pm 0,0300$ ml
Average value:	19,9993 ml
Deviation:	0,0052 ml
Precision (CV):	0,0261 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	11.07.2024
Balance:	Sartorius - LE244S-OCE

