

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 200

Volumetric pipettes, 20 ml, AS, conf. cert.

Batch: 28-19


Serial:

Measuring Nr. 31772

Batch size:

1168

Volume:	20,000 ml
Tolerance:	$\pm 0,0300$ ml
Average value:	19,9983 ml
Deviation:	0,0060 ml
Precision (CV):	0,0302 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

22.07.2019

Balance:

Sartorius -
LE244S-OCE

DE-M