

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 200

Volumetric pipettes, 20 ml, AS, conf. cert.

Batch: 50-19


Serial:

Measuring Nr. 35293

Batch size:

2387

Volume:	20,0000 ml
Tolerance:	$\pm 0,0300$ ml
Average value:	19,9947 ml
Deviation:	0,0065 ml
Precision (CV):	0,0325 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	18.12.2019
Balance:	Sartorius - MSE224S-0CE -DU

DE-M