

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 200

**Volumetric pipettes, 20 ml, AS, conf. cert.**

**Batch:** 23-20


**Serial:**

**Measuring Nr.** 39201

**Batch size:**

1111

<b>Volume:</b>	20,000 ml
<b>Tolerance:</b>	$\pm 0,0300$ ml
<b>Average value:</b>	20,0018 ml
<b>Deviation:</b>	0,0083 ml
<b>Precision (CV):</b>	0,0416 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

24.07.2020

**Balance:**

Sartorius -  
MSE224S-0CE  
-DU

DE-M