

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 400

Volumetric pipettes, 40 ml, AS, conf. cert.

Batch: 49-17


Serial:

Measuring Nr. 16764

Batch size:

156

Volume:	40,0000 ml
Tolerance:	$\pm 0,0500$ ml
Average value:	40,0072 ml
Deviation:	0,0066 ml
Precision (CV):	0,0164 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

21.12.2017

Balance:

Sartorius -
LE244S-OCE

DE-M