

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 400

Volumetric pipettes, 40 ml, AS, conf. cert.

Batch: 21-22


Serial:

Measuring Nr. 56491

Batch size:

145

| | |
|------------------------|-----------------|
| Volume: | 40,000 ml |
| Tolerance: | $\pm 0,0500$ ml |
| Average value: | 40,0012 ml |
| Deviation: | 0,0086 ml |
| Precision (CV): | 0,0215 % |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

03.06.2022

Balance:

Sartorius -
MSE224S-0CE
-DU

DE-M