

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 500 USP

Vol. pipette, 50 ml, USP, AS, conf. cert

Batch: 19-20


Serial:

Measuring Nr. 37936

Batch size:

60

Volume:	50,0000 ml
Tolerance:	± 0,0500 ml
Average value:	49,9945 ml
Deviation:	0,0069 ml
Precision (CV):	0,0137 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

7

Date:

14.05.2020

Balance:

Sartorius -
MSE224S-0CE
-DU

DE-M