

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 750

Volumetric pipettes, 75 ml, AS, conf. cert


Batch: 15-19

Serial:

Measuring Nr. 29689

Batch size: 60

Volume:	75,0000 ml
Tolerance:	$\pm 0,0500$ ml
Average value:	75,0096 ml
Deviation:	0,0070 ml
Precision (CV):	0,0094 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 7

Date: 07.05.2019

Balance: Sartorius -
LE244S-OCE

DE-M