

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 900 BL

Volumetric pipettes, 100 ml, blue grad., AS, conf. cert.


Batch: 38-17

Serial:

Measuring Nr. 15236

Batch size: 180

Volume:	100,000 ml
Tolerance:	$\pm 0,0800$ ml
Average value:	100,0250 ml
Deviation:	0,0089 ml
Precision (CV):	0,0089 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 4

Date: 20.10.2017

Balance: Sartorius -
LE244S-OCE

DE-M