

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 900 BL

**Volumetric pipettes, 100 ml, blue grad., AS, conf. cert.**


**Batch:** 13-18

**Serial:**

**Measuring Nr.** 20286

**Batch size:** 275

<b>Volume:</b>	100,0000 ml
<b>Tolerance:</b>	$\pm 0,0800$ ml
<b>Average value:</b>	100,0108 ml
<b>Deviation:</b>	0,0196 ml
<b>Precision (CV):</b>	0,0196 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:** 44

**Date:** 03.05.2018

**Balance:** Sartorius -  
LE244S-OCE

DE-M