

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 900

Volumetric pipettes, 100 ml, AS, conf.cert.


Batch: 35-20

Serial:

Measuring Nr. 41717

Batch size: 495

Volume:	100,000 ml
Tolerance:	$\pm 0,0800$ ml
Average value:	100,0243 ml
Deviation:	0,0239 ml
Precision (CV):	0,0239 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	27.10.2020
Balance:	Sartorius - LE244S-OCE

DE-M