

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 920

Volumetric pipettes, 200 ml, AS, KB


Batch: 11-18

Serial:

Measuring Nr. 18947

Batch size: 25

Volume:	200,000 ml
Tolerance:	$\pm 0,0800$ ml
Average value:	200,0190 ml
Deviation:	0,0248 ml
Precision (CV):	0,0124 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 4

Date: 19.03.2018

Balance: Sartorius - LA
1200 S

DE-M